


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 9962 Accredited to ISO/IEC 17025:2017	Atlas Copco Ltd trading as Atlas Copco Tools and Industrial Assembly Solutions	
	Issue No: 012 Issue date: 14 July 2023	
	Swallowdale Lane Hemel Hempstead Hertfordshire HP2 7EA	Contact: Mr Nicolas Van Zyl Telephone: +44 (0) 1442 261202 Email: lab.uk@atlascopco.com Website: www.atlascopco.co.uk
Calibration performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Innovation Drive Coven Wolverhampton WV9 5GA Local contact Mr Nicolas Van Zyl	TORQUE	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer Contact Mr Nicolas Van Zyl	TORQUE	S



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Atlas Copco Ltd trading as
Atlas Copco Tools and Industrial Assembly Solutions**

Issue No: 012 Issue date: 14 July 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
TORQUE				P, S
Hand torque tools (excluding torque screwdrivers)	To BS EN ISO 6789-2:2017 1 N·m to 1500 N·m	1.5 %	The uncertainties quoted are for both the application of the calibration torque and the characteristics of the device being calibrated. Calibration results may also be given in imperial units.	
Mechanical and Electronic Torque Calibration Equipment	To BS EN 7882:2017 0.2 N·m to 1400 N·m	1.0 %	Calibrated statically using an un-supported beam and masses.	P
Mechanical and Electronic Torque Calibration Equipment	To BS EN 7882:2017 0.2 N·m to 3000 N·m	1.0 %	Calibrated statically by comparison to a reference transducer.	P, S
Torque measuring device angle parameters	To VDI/VDE2648-1 0° to 360°	1.0°	Calibrated statically by comparison to a reference angle encoder.	P, S
Electrically powered and controlled Torque devices	in House method 0.2 N·m to 2000 N·m	2.5%	Static calibration against reference torque transducer. This does not include Pneumatic and Hydraulic powered devices.	P, S
END				



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Atlas Copco Ltd trading as
Atlas Copco Tools and Industrial Assembly Solutions**

Issue No: 012 Issue date: 14 July 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$